

Lovibond® Colour Measurement Tintometer® Group



AOCS-Tintometer Filter Set (Fx)

Part Number: 166900



Conformance filters are assigned a nominal value and the company always endeavors to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

AOCS-Tintometer Filter Set (Fx)

A set of three filters for testing the PFXi 880, 950 or 995 in AOCS-Tintometer scale.

Technical Data

Glass Standard Values	Approximate Values
	- 0.5R 1.0Y,
	- 3.7R 13.0Y,
	- 6.6R 50Y

Filters	3
---------	---

Delivery Scope

- 0.5R 1.0Y
- 3.7R 13.0Y
- 6.6R 50Y
- Certificate of Conformity